Document Reference: PH-DOC-078

Document Issue Number: 07

Document Issue Date: 08/03/2024

Page 1 of 2



Tel: +44 (0)33 3320 8550

EU Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

| MANUFACTURER | | |
|-----------------------|--|-----------------|
| Name of Company | Address | SRN |
| Polyco Healthline Ltd | South Fen Road, Bourne, Lincolnshire, PE10 0DN. UK | GB-MF-000015289 |

| AUTHORIZED REPRESENTATIVE | | | | | |
|---|---|-----------------|-------------------------------------|--|--|
| Name of Company | Address | SRN | Phone/email | | |
| International Associates Auditing & Certification Limited | The Black Church, St Mary's Place, Dublin 7, D07 P4AX Ireland | IE-AR-000002248 | +353 16971561 EUAR@ie.ia-net.com | | |

| Product Name | Code / Catalogue Number | Basic UDI-DI |
|---|-------------------------|-------------------|
| Black Nitrile Powder Free Examination Glove | GL897 | 5024951GNPFNS00YA |
| Intended Purpose | Photo | |
| Examination gloves for non-invasive use to prevent contact with skin, bodily fluids, and chemicals. Invasive with respect to natural body orifices. | | |

| RISK CLASS FOR MEDICAL DEVICES | | | | |
|--------------------------------|--------|-----------------------------------|---|--|
| Device Classific | cation | Common Specifications / Standards | | |
| Class: | I | EN455-1:2020 EN455-2:2015 | Medical gloves for single use: Freedom from holes Medical gloves for single use: Physical properties | |
| Rule: | 5 | EN455-3:2015 EN455-4:2009 | Medical gloves for single use: Biological evaluation Medical gloves for single use: Shelf-life determination | |

| RISK CATEGORY OF PERSONAL PROTECTIVE EQUIPMENT AND PERFORMANCE LEVELS | | | |
|---|---------------------------------------|----------------------------|--|
| Product Risk Category | | | |
| Personal Protective Equipment (PPE) Category III | | | |
| Standard | Performance Levels | | |
| EN ISO 374-1:2016 | Type C - K | | |
| EN ISO 374-5:2016 | Protection against bacteria and fungi | Protection against viruses | |
| | Pass | Pass | |

Polyco Healthline Ltd declares that the above-mentioned product meets the provision of the following EU legislation:

Head Office & Distribution Centres: South Fen Road, Bourne, Lincolnshire, PE10 0DN, UK

Document Reference: PH-DOC-078

Document Issue Number: 07

Document Issue Date: 08/03/2024

Page 2 of 2



- Medical Devices Regulation (EU) 2017/745
- Is classed as Cat III PPE in conformity with the provisions of Regulation (EU) 2016/425 on personal protective equipment and with National Standards transposing harmonised standards EN ISO 374-1:2016, EN ISO 374-5:2016, and EN 21420:2020.
- is identical to the personal protective equipment which is the subject of EU Type-Examination Certificate number 2777/11719-04/E00-00, issued by SATRA Technology Europe Limited, Bracetown Business Park, Clonee, Dublin, D15 YN2P, Republic of Ireland (Notified Body number 2777), according to Annex V (Module B) of Regulation (EU) 2016/425.
- is subject to the conformity assessment procedure Module C2 set out in Annex VII of Regulation (EU) 2016/425, under the surveillance of the notified body SATRA Technology Europe Limited, Bracetown Business Park, Clonee, Dublin, D15 YN2P, Republic of Ireland (Notified Body number 2777).

COMPANY REPRESENTATIVE: David Langridge

TITLE: Head of Technical SIGNATURE: Designing

PLACE: Bourne, UK ISSUE DATE: 08th March 2024